
Midmark M11 - v1.0.5
Total Cycles: 3804

Name: RITTER2
Sterilizer ID: V2313228
07/22/2024 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.1 F 0.0
2:00 123.6 F 0.8
4:00 131.6 F 2.9
6:00 161.4 F 6.1
8:00 185.2 F 9.7
10:00 198.9 F 13.0
12:00 211.4 F 5.4
14:00 223.4 F 9.1
16:00 234.7 F 13.4
18:00 235.3 F 10.0
20:00 239.3 F 9.7
22:00 246.0 F 13.7
24:00 256.6 F 19.5
26:00 266.4 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.9 F 29.8
2:00 271.6 F 29.0
3:00 272.7 F 29.9
4:00 272.0 F 29.3

Min 271.5 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 195.4 F 0.0
10:00 202.0 F 0.0
15:00 203.4 F 0.0
20:00 200.4 F 0.0
25:00 204.5 F 0.0
30:00 207.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 27:42
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:06:26
CYCLE COMPLETE

07/22/2024 11:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____