
Midmark M11 - v1.0.0
Total Cycles: 2919

Name: RITTER 1
Sterilizer ID: V2178198
02/17/2023 12:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.2 F 0.0
2:00 174.2 F 2.7
4:00 181.3 F 6.7
6:00 204.7 F 11.8
8:00 223.5 F 6.9
10:00 237.4 F 12.6
12:00 236.5 F 9.9
14:00 236.8 F 8.3
16:00 248.4 F 14.7
18:00 260.6 F 21.4
20:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.4 F 29.2
2:00 271.2 F 28.5
3:00 272.4 F 29.4
4:00 271.6 F 28.7

Min 271.1 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 230.7 F 0.0
10:00 221.0 F 0.0
15:00 220.0 F 0.0
20:00 218.0 F 0.0
25:00 217.0 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

02/17/2023 01:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____