
Midmark M11 - v1.0.0
Total Cycles: 3244

Name: RITTER 1
Sterilizer ID: V2178198
05/16/2023 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.7 F 0.0
2:00 194.9 F 2.4
4:00 208.7 F 8.0
6:00 226.6 F 11.0
8:00 234.3 F 10.2
10:00 233.8 F 8.0
12:00 240.7 F 10.6
14:00 247.9 F 14.3
16:00 260.8 F 21.5
18:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.3 F 29.0
2:00 271.5 F 28.5
3:00 272.4 F 29.0
4:00 271.4 F 28.3

Min 271.3 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 229.0 F 0.0
10:00 236.7 F 0.0
15:00 238.0 F 0.0
20:00 230.7 F 0.0
25:00 220.9 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 18:26
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:57:42
CYCLE COMPLETE

05/16/2023 03:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____