
Midmark M11 - v1.0.0
Total Cycles: 3149

Name: RITTER 1
Sterilizer ID: V2178198
04/24/2023 07:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.1 F 0.0
2:00 79.4 F 0.0
4:00 86.7 F 0.0
6:00 107.7 F 0.0
8:00 157.2 F 0.0
10:00 187.6 F 1.5
12:00 203.1 F 5.0
14:00 215.0 F 8.1
16:00 226.0 F 11.7
18:00 225.7 F 6.4
20:00 237.5 F 11.3
22:00 234.0 F 8.2
24:00 242.6 F 11.3
26:00 245.8 F 13.0
28:00 258.0 F 19.5
30:00 268.8 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.3 F 28.5
2:00 272.3 F 28.9
3:00 271.8 F 28.2
4:00 272.7 F 29.0

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 193.5 F 0.0
10:00 210.9 F 0.0
15:00 213.3 F 0.0
20:00 216.7 F 0.0
25:00 219.3 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:06
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:09
CYCLE COMPLETE

04/24/2023 08:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____