
Midmark M11 - v1.0.5
Total Cycles: 1186

Name: RITTER2
Sterilizer ID: V2313228
06/13/2022 08:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.4 F 0.0
2:00 166.4 F 0.0
4:00 167.0 F 0.3
6:00 195.9 F 5.7
8:00 217.9 F 11.7
10:00 228.9 F 8.1
12:00 239.0 F 11.2
14:00 243.7 F 12.7
16:00 243.2 F 12.0
18:00 256.7 F 19.2
20:00 268.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.8 F 29.2
2:00 271.8 F 28.9
3:00 272.5 F 29.0
4:00 271.9 F 28.9

Min 271.4 F 28.4
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 227.0 F 0.0
5:00 233.9 F 0.0
10:00 226.8 F 0.0
15:00 222.7 F 0.0
20:00 219.9 F 0.0
Dry Cycle Aborted

FILLING: 1:03
HEATING: 21:09
STERILIZING: 4:00
VENTING: 2:41
DRYING: 21:45
TOTAL CYCLE: 00:50:38

06/13/2022 09:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____