

-----  
Midmark M11 - v1.0.0  
Total Cycles: 912

Name: RITTER 1  
Sterilizer ID: V2178198  
06/17/2021 02:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 152.7 F 0.0  
2:00 154.6 F 1.7  
4:00 162.2 F 4.5  
6:00 185.9 F 8.2  
8:00 205.3 F 12.7  
10:00 223.0 F 6.6  
12:00 230.3 F 11.7  
14:00 233.0 F 8.5  
16:00 242.7 F 11.2  
18:00 244.2 F 12.6  
20:00 256.3 F 18.9  
22:00 267.0 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.5 F 29.2  
2:00 271.8 F 28.9  
3:00 271.9 F 28.9  
4:00 272.3 F 29.1

Min 271.0 F 28.4  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.9 F 0.0  
5:00 233.1 F 0.0  
10:00 241.0 F 0.0  
15:00 234.5 F 0.0  
20:00 225.7 F 0.0  
25:00 227.2 F 0.0  
30:00 228.8 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 23:29  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:38  
CYCLE COMPLETE

06/17/2021 03:25 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_