
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000401
16:21 25/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 3:01
STERILIZING 3:59
136.0 C 226kPa 3:59
136.4 C 228kPa 4:29
136.3 C 233kPa 4:59
136.3 C 233kPa 5:29
136.4 C 233kPa 5:59
136.2 C 230kPa 6:29
136.4 C 233kPa 6:59
136.4 C 233kPa 7:29
136.4 C 233kPa 7:29
VENTING 7:29
AIR DRYING 8:24
CYCLE COMPLETE 8:25

Digital Signature #
408626BD908B3F92
