
Midmark M11 - v1.0.5
Total Cycles: 3390

Name: RITTER2
Sterilizer ID: V2313228
02/16/2024 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.5 F 0.0
2:00 151.2 F 0.8
4:00 157.0 F 3.5
6:00 181.7 F 7.3
8:00 201.4 F 11.6
10:00 216.6 F 5.7
12:00 226.9 F 10.3
14:00 231.2 F 8.3
16:00 241.6 F 12.9
18:00 241.6 F 11.8
20:00 253.9 F 18.0
22:00 264.8 F 24.6
24:00 271.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.5 F 29.9
2:00 271.8 F 29.3
3:00 271.3 F 29.5
4:00 272.5 F 29.9

Min 271.1 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 206.6 F 0.0
10:00 198.5 F 0.0
15:00 196.0 F 0.0
20:00 193.6 F 0.0
25:00 191.7 F 0.0
30:00 189.9 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:40
CYCLE COMPLETE

02/16/2024 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____