

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2237

Name: RITTER 1  
Sterilizer ID: V2178198  
08/24/2022 08:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.8 F 0.0  
2:00 163.2 F 0.1  
4:00 171.5 F 0.6  
6:00 196.3 F 3.8  
8:00 217.6 F 8.3  
10:00 231.7 F 13.2  
12:00 233.9 F 9.2  
14:00 240.0 F 10.1  
16:00 244.2 F 12.6  
18:00 242.1 F 11.2  
20:00 255.1 F 17.9  
22:00 266.4 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.1 F 28.3  
2:00 272.6 F 28.9  
3:00 271.5 F 28.2  
4:00 272.1 F 28.4

Min 271.4 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.2 F 0.0  
5:00 227.6 F 0.0  
10:00 238.9 F 0.0  
15:00 221.8 F 0.0  
20:00 215.1 F 0.0  
25:00 216.3 F 0.0  
30:00 229.9 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 23:33  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:33  
CYCLE COMPLETE

08/24/2022 10:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_