

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.9uS / 0.5ppm  
CYCLE NUMBER 000834  
12:33 16/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:36  
PRESSURIZING 3:37  
STERILIZING 4:58  
136.4 C 222kPa 4:58  
136.4 C 224kPa 5:28  
136.4 C 224kPa 5:58  
136.3 C 223kPa 6:28  
136.4 C 225kPa 6:58  
136.3 C 224kPa 7:28  
136.2 C 221kPa 7:58  
136.2 C 223kPa 8:28  
136.2 C 223kPa 8:28  
VENTING 8:28  
AIR DRYING 9:23  
CYCLE COMPLETE 9:24

-----  
Digital Signature #  
826F90285AB12C29  
-----