
Midmark M11 - v1.0.0
Total Cycles: 3506

Name: RITTER 1
Sterilizer ID: V2178198
07/10/2023 11:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.5 F 0.0
2:00 198.1 F 2.9
4:00 203.2 F 7.6
6:00 221.7 F 12.9
8:00 228.5 F 8.3
10:00 241.3 F 12.7
12:00 240.7 F 11.4
14:00 238.8 F 9.9
16:00 252.5 F 16.6
18:00 264.2 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.1 F 28.6
2:00 272.5 F 29.1
3:00 271.6 F 28.3
4:00 272.4 F 28.8

Min 271.5 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 218.7 F 0.0
10:00 220.8 F 0.0
15:00 224.5 F 0.0
20:00 225.5 F 0.0
25:00 225.6 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

07/10/2023 12:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____