

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3411

Name: RITTER2  
Sterilizer ID: V2313228  
02/26/2024 11:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.9 F 0.0  
2:00 143.3 F 0.3  
4:00 142.4 F 2.2  
6:00 167.6 F 5.9  
8:00 188.4 F 10.0  
10:00 221.8 F 7.6  
12:00 223.1 F 8.8  
14:00 234.8 F 13.5  
16:00 236.2 F 10.3  
18:00 238.9 F 8.8  
20:00 246.5 F 14.1  
22:00 257.5 F 20.1  
24:00 268.0 F 26.7  
26:00 272.5 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.7  
1:00 272.4 F 29.8  
2:00 272.7 F 30.1  
3:00 271.8 F 30.1  
4:00 272.1 F 29.7

Min 271.7 F 29.6  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.8 F 0.0  
5:00 221.1 F 0.0  
10:00 222.7 F 0.0  
15:00 226.4 F 0.0  
20:00 228.9 F 0.0  
25:00 227.9 F 0.0  
30:00 229.4 F 0.0

Drying Complete

FILLING: 0:58  
HEATING: 26:10  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:04:41  
CYCLE COMPLETE

02/26/2024 12:33 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_