
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001456
8:55 09/29/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:14
PRESSURIZING 7:39
STERILIZING 9:28
136.0 C 220kPa 9:28
136.4 C 220kPa 9:58
136.5 C 221kPa 10:28
136.4 C 222kPa 10:59
136.4 C 220kPa 11:29
136.2 C 220kPa 11:59
136.4 C 222kPa 12:29
136.0 C 220kPa 12:58
VENTING 12:59
AIR DRYING 14:07
CYCLE COMPLETE 44:08

Digital Signature #
4E801385B7B2A6F0
