
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005649
8:19 02/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:05
STERILIZING 4:18
136.4 C 222kPa 4:18
135.8 C 220kPa 4:48
136.1 C 222kPa 5:18
136.2 C 222kPa 5:48
136.4 C 223kPa 6:18
136.4 C 220kPa 6:48
136.4 C 223kPa 7:18
136.2 C 221kPa 7:48
136.2 C 221kPa 7:48
VENTING 7:48
AIR DRYING 8:36
CYCLE COMPLETE 8:41

Digital Signature #
17DD4BAC63033548
