
Midmark M11 - v1.0.0
Total Cycles: 3446

Name: RITTER 1
Sterilizer ID: V2178198
06/26/2023 01:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.0 F 0.0
2:00 189.7 F 3.4
4:00 198.8 F 7.7
6:00 216.8 F 13.4
8:00 227.1 F 8.2
10:00 240.6 F 13.0
12:00 240.6 F 11.4
14:00 239.2 F 10.1
16:00 253.3 F 17.1
18:00 265.0 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 29.0
2:00 271.6 F 28.5
3:00 272.6 F 29.1
4:00 271.8 F 28.5

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 230.7 F 0.0
10:00 227.1 F 0.0
15:00 228.9 F 0.0
20:00 228.1 F 0.0
25:00 229.6 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:47
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:41
CYCLE COMPLETE

06/26/2023 02:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____