
Midmark M11 - v1.0.0
Total Cycles: 1294

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2021 06:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

| mm:ss | Degrees | PSI |
|-------|---------|------|
| 0:00 | 70.8 F | 0.0 |
| 2:00 | 73.5 F | 0.0 |
| 4:00 | 75.4 F | 0.0 |
| 6:00 | 94.0 F | 0.1 |
| 8:00 | 129.1 F | 0.1 |
| 10:00 | 165.6 F | 0.2 |
| 12:00 | 188.3 F | 0.3 |
| 14:00 | 213.2 F | 2.0 |
| 16:00 | 221.4 F | 4.9 |
| 18:00 | 228.2 F | 7.5 |
| 20:00 | 235.7 F | 10.7 |
| 22:00 | 237.7 F | 9.0 |
| 24:00 | 242.8 F | 11.4 |
| 26:00 | 237.4 F | 9.0 |
| 28:00 | 241.0 F | 10.1 |
| 30:00 | 247.9 F | 13.8 |
| 32:00 | 259.1 F | 19.8 |
| 34:00 | 269.1 F | 26.2 |

STERILIZING

| mm:ss | Degrees | PSI |
|-------|---------|------|
| 0:00 | 271.4 F | 27.8 |
| 1:00 | 272.6 F | 28.5 |
| 2:00 | 271.5 F | 27.7 |
| 3:00 | 272.6 F | 28.6 |
| 4:00 | 271.8 F | 27.9 |

Min 271.4 F 27.7
Max 272.6 F 28.6

VENTING CHAMBER

DRYING

Drying Start

| mm:ss | Degrees | PSI |
|-------|---------|-----|
| 0:00 | 176.9 F | 0.0 |
| 5:00 | 220.8 F | 0.0 |
| 10:00 | 220.9 F | 0.0 |
| 15:00 | 217.9 F | 0.0 |
| 20:00 | 211.4 F | 0.0 |
| 25:00 | 210.6 F | 0.0 |
| 30:00 | 210.8 F | 0.0 |

Drying Complete

FILLING: 0:53
HEATING: 35:10
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:13:22
CYCLE COMPLETE

10/25/2021 07:59 AM
ACCEPTED: X
REJECTED:
Operator: U01