

-----  
Midmark M11 - v1.0.0  
Total Cycles: 652

Name: MIDMARK  
Sterilizer ID: V2178198  
02/08/2021 07:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 71.8 F 0.0  
4:00 76.7 F 0.1  
6:00 108.6 F 0.1  
8:00 144.1 F 0.2  
10:00 180.0 F 0.9  
12:00 207.1 F 4.0  
14:00 214.0 F 7.0  
16:00 224.9 F 10.4  
18:00 229.9 F 9.0  
20:00 235.0 F 9.2  
22:00 240.4 F 11.0  
24:00 242.5 F 11.5  
26:00 238.6 F 9.4  
28:00 252.6 F 16.2  
30:00 263.8 F 22.7  
32:00 271.5 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.8  
1:00 272.4 F 28.3  
2:00 271.8 F 28.2  
3:00 272.1 F 28.0  
4:00 272.4 F 28.6  
  
Min 271.3 F 27.7  
Max 272.7 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.7 F 0.0  
5:00 240.7 F 0.0  
10:00 230.7 F 0.0  
15:00 227.6 F 0.0  
20:00 209.0 F 0.0  
25:00 197.3 F 0.0  
30:00 198.8 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 32:06  
STERILIZING: 4:00  
VENTING: 2:51  
DRYING: 30:00  
TOTAL CYCLE: 01:09:46  
CYCLE COMPLETE

02/08/2021 08:47 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01