
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001716
10:59 21/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:40
STERILIZING 3:43
136.4 C 225kPa 3:43
136.2 C 224kPa 4:13
136.2 C 223kPa 4:43
136.4 C 225kPa 5:13
136.4 C 226kPa 5:43
136.4 C 224kPa 6:13
136.3 C 224kPa 6:43
136.4 C 221kPa 7:13
136.4 C 221kPa 7:13
VENTING 7:13
AIR DRYING 8:04
CYCLE COMPLETE 8:20

Digital Signature #
BE9935069F208ADA
