
Midmark M11 - v1.0.0
Total Cycles: 4891

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2024 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.9 F 0.0
2:00 158.5 F 0.1
4:00 175.5 F 0.6
6:00 203.0 F 2.8
8:00 210.8 F 6.8
10:00 223.5 F 10.8
12:00 228.4 F 6.4
14:00 236.7 F 10.9
16:00 234.0 F 8.0
18:00 243.5 F 11.7
20:00 246.1 F 13.0
22:00 257.6 F 19.1
24:00 267.7 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.1 F 28.6
2:00 272.7 F 28.9
3:00 272.4 F 28.6
4:00 272.1 F 28.3

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 228.9 F 0.0
10:00 237.2 F 0.0
15:00 230.1 F 0.0
20:00 224.6 F 0.0
25:00 223.5 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 25:37
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:04:12
CYCLE COMPLETE

09/11/2024 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____