
Midmark M11 - v1.0.0
Total Cycles: 5484

Name: RITTER 1
Sterilizer ID: V2178198
04/07/2025 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.3 F 0.0
2:00 122.5 F 1.5
4:00 134.8 F 4.0
6:00 160.5 F 7.2
8:00 189.1 F 10.9
10:00 211.8 F 4.9
12:00 213.4 F 7.3
14:00 225.9 F 11.6
16:00 228.7 F 7.4
18:00 241.7 F 12.9
20:00 239.8 F 10.8
22:00 252.3 F 17.0
24:00 263.4 F 23.4
26:00 272.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.5 F 29.6
2:00 272.6 F 29.3
3:00 271.7 F 28.8
4:00 271.1 F 28.7

Min 271.1 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 198.9 F 0.0
10:00 188.6 F 0.0
15:00 187.0 F 0.0
20:00 187.6 F 0.0
25:00 195.0 F 0.0
30:00 209.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:15
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:04:38
CYCLE COMPLETE

04/07/2025 02:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____