
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000004
12:08 30/11/2019
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 2:57
CYCLE FAULT 15
NOT STERILE
END TIME 2:57

Digital Signature #
A1EDD2389F43AB67
