
Midmark M11 - v1.0.0
Total Cycles: 4344

Name: RITTER 1
Sterilizer ID: V2178198
02/29/2024 09:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.9 F 0.0
2:00 96.9 F 0.2
4:00 85.7 F 1.1
6:00 110.0 F 3.2
8:00 139.0 F 5.9
10:00 164.7 F 9.3
12:00 193.0 F 13.1
14:00 213.8 F 4.0
16:00 218.9 F 7.2
18:00 227.3 F 10.6
20:00 231.1 F 7.7
22:00 238.6 F 10.4
24:00 235.7 F 8.2
26:00 247.3 F 13.7
28:00 258.4 F 19.7
30:00 268.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.5 F 28.8
2:00 271.4 F 28.0
3:00 272.1 F 28.8
4:00 272.3 F 28.5

Min 271.1 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 230.1 F 0.0
10:00 231.1 F 0.0
15:00 226.8 F 0.0
20:00 222.7 F 0.0
25:00 234.5 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:17
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:09:30
CYCLE COMPLETE

02/29/2024 10:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____