
Midmark M11 - v1.0.5
Total Cycles: 3441

Name: RITTER2
Sterilizer ID: V2313228
03/07/2024 03:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.8 F 0.0
2:00 117.5 F 0.4
4:00 139.1 F 2.6
6:00 167.9 F 5.8
8:00 183.7 F 9.3
10:00 201.0 F 13.3
12:00 215.2 F 6.7
14:00 224.5 F 10.9
16:00 232.0 F 7.6
18:00 241.3 F 13.1
20:00 242.3 F 11.9
22:00 254.0 F 18.1
24:00 265.2 F 24.6
26:00 270.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.1 F 29.5
2:00 272.7 F 29.7
3:00 272.4 F 30.0
4:00 271.7 F 29.8

Min 271.6 F 29.4
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.5 F 0.0
5:00 229.8 F 0.0
10:00 211.4 F 0.0
15:00 212.2 F 0.0
20:00 194.1 F 0.0
25:00 187.9 F 0.0
30:00 184.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 26:38
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:05:26
CYCLE COMPLETE

03/07/2024 04:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____