
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003893
14:44 27/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:35
STERILIZING 3:41
136.4 C 222kPa 3:41
136.3 C 224kPa 4:11
136.1 C 223kPa 4:41
136.3 C 225kPa 5:11
136.4 C 225kPa 5:41
136.4 C 224kPa 6:11
136.2 C 218kPa 6:41
136.4 C 223kPa 7:11
136.4 C 223kPa 7:11
VENTING 7:12
AIR DRYING 8:03
CYCLE COMPLETE 8:16

Digital Signature #
C072B149323F4052
