

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2750

Name: RITTER 1  
Sterilizer ID: V2178198  
01/09/2023 11:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.5 F 0.0  
2:00 162.0 F 0.0  
4:00 168.0 F 0.0  
6:00 200.0 F 4.0  
8:00 207.6 F 8.9  
10:00 224.2 F 11.8  
12:00 230.5 F 8.8  
14:00 240.2 F 11.8  
16:00 240.8 F 11.4  
18:00 238.3 F 9.7  
20:00 251.7 F 16.2  
22:00 263.1 F 22.8  
24:00 272.5 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.1  
2:00 272.0 F 28.4  
3:00 271.5 F 28.7  
4:00 272.6 F 29.1

Min 271.3 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.4 F 0.0  
5:00 218.0 F 0.0  
10:00 233.1 F 0.0  
15:00 228.9 F 0.0  
20:00 230.8 F 0.0  
25:00 229.2 F 0.0  
30:00 227.4 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 24:17  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:02:18  
CYCLE COMPLETE

01/09/2023 12:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_