

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1745

Name: RITTER2  
Sterilizer ID: V2313228  
11/17/2022 03:52 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
2:00 180.0 F 2.4  
4:00 202.5 F 7.6  
6:00 220.1 F 13.3  
8:00 230.3 F 9.5  
10:00 234.8 F 9.0  
12:00 241.6 F 11.5  
14:00 246.9 F 14.2  
16:00 260.0 F 21.4  
18:00 271.7 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.3 F 29.4  
2:00 271.8 F 29.4  
3:00 272.2 F 29.3  
4:00 271.7 F 29.3  
  
Min 271.5 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
5:00 222.3 F 0.0  
10:00 225.6 F 0.0  
15:00 225.4 F 0.0  
20:00 224.0 F 0.0  
25:00 222.6 F 0.0  
30:00 223.0 F 0.0  
  
Drying Complete

FILLING: 1:47  
HEATING: 18:33  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:33  
CYCLE COMPLETE

11/17/2022 04:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_