
Midmark M11 - v1.0.5
Total Cycles: 460

Name: RITTER2
Sterilizer ID: V2313228
10/19/2021 10:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.7 F 0.0
2:00 193.8 F 3.4
4:00 202.7 F 9.8
6:00 224.5 F 7.5
8:00 233.8 F 11.9
10:00 239.7 F 12.7
12:00 242.4 F 12.2
14:00 256.2 F 19.4
16:00 268.4 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 30.3
1:00 272.6 F 30.7
2:00 272.4 F 30.4
3:00 272.5 F 30.2
4:00 272.3 F 30.4

Min 272.0 F 30.2
Max 273.0 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 234.8 F 0.0
10:00 223.0 F 0.0
15:00 219.1 F 0.0
20:00 216.5 F 0.0
25:00 216.0 F 0.0
Dry Cycle Aborted

FILLING: 2:33
HEATING: 17:54
STERILIZING: 4:00
VENTING: 3:00
DRYING: 25:24
TOTAL CYCLE: 00:52:51

10/19/2021 11:16 AM
ACCEPTED: X
REJECTED:
Operator: U01