
Midmark M11 - v1.0.5
Total Cycles: 4014

Name: RITTER2
Sterilizer ID: V2313228
10/09/2024 02:31 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.0 F 0.0
2:00 101.9 F 0.0
4:00 109.2 F 0.0
6:00 144.3 F 0.2
8:00 170.3 F 2.7
10:00 200.6 F 6.0
12:00 212.8 F 9.5
14:00 224.3 F 13.3
16:00 227.6 F 8.4
18:00 238.5 F 13.0
20:00 237.2 F 10.5
22:00 235.6 F 8.8
24:00 247.9 F 14.8
26:00 258.7 F 20.8
28:00 268.5 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.4 F 29.6
2:00 271.8 F 29.7
3:00 272.1 F 29.4
4:00 271.8 F 29.8

Min 271.3 F 29.1
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 216.7 F 0.0
10:00 204.2 F 0.0
15:00 217.3 F 0.0
20:00 219.5 F 0.0
25:00 228.7 F 0.0
30:00 233.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:13
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:07:55
CYCLE COMPLETE

10/09/2024 03:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____