
Midmark M11 - v1.0.0
Total Cycles: 616

Name: MIDMARK
Sterilizer ID: V2178198
01/21/2021 08:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 77.7 F 0.0
4:00 99.9 F 0.0
6:00 160.7 F 0.1
8:00 179.2 F 0.3
10:00 194.9 F 2.9
12:00 203.9 F 6.2
14:00 217.7 F 9.9
16:00 227.1 F 9.1
18:00 231.1 F 8.9
20:00 240.3 F 11.7
22:00 240.6 F 11.4
24:00 238.0 F 9.6
26:00 252.0 F 16.5
28:00 263.7 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.0 F 28.5
2:00 272.5 F 29.0
3:00 272.4 F 29.2
4:00 271.6 F 28.4

Min 271.6 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 164.5 F 0.0
5:00 222.1 F 0.0
10:00 232.8 F 0.0
15:00 227.7 F 0.0
20:00 228.0 F 0.0
25:00 229.6 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 29:60
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 01:07:56
CYCLE COMPLETE

01/21/2021 09:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____