

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2719

Name: RITTER 1  
Sterilizer ID: V2178198  
12/22/2022 01:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.4 F 0.0  
2:00 180.0 F 2.4  
4:00 190.9 F 6.4  
6:00 221.9 F 11.9  
8:00 227.0 F 6.4  
10:00 236.8 F 11.3  
12:00 236.2 F 8.0  
14:00 244.6 F 11.5  
16:00 245.4 F 12.7  
18:00 258.2 F 19.4  
20:00 269.2 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.4 F 29.0  
2:00 272.0 F 28.5  
3:00 271.2 F 27.9  
4:00 271.8 F 28.7

Min 271.0 F 27.9  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.6 F 0.0  
5:00 225.0 F 0.0  
10:00 241.5 F 0.0  
15:00 234.0 F 0.0  
20:00 226.3 F 0.0  
25:00 225.4 F 0.0  
30:00 233.6 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 21:08  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:49  
CYCLE COMPLETE

12/22/2022 02:02 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_