
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006463
8:06 16/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:51
PRESSURIZING 2:54
STERILIZING 3:58
136.4 C 221kPa 3:58
136.2 C 222kPa 4:28
136.4 C 222kPa 4:58
136.2 C 219kPa 5:28
136.4 C 219kPa 5:58
136.4 C 220kPa 6:28
136.4 C 221kPa 6:58
136.4 C 222kPa 7:28
136.4 C 222kPa 7:28
VENTING 7:28
AIR DRYING 8:17
CYCLE COMPLETE 8:20

Digital Signature #
D4AAC559BFE83BAD
