
Midmark M11 - v1.0.0
Total Cycles: 3924

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2023 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.0 F 0.0
2:00 177.6 F 2.9
4:00 185.1 F 7.0
6:00 209.5 F 12.3
8:00 221.0 F 6.6
10:00 232.1 F 11.8
12:00 233.8 F 8.8
14:00 239.8 F 10.5
16:00 246.0 F 13.4
18:00 257.8 F 19.7
20:00 268.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.2 F 29.1
2:00 272.8 F 29.2
3:00 272.3 F 29.0
4:00 272.5 F 29.5

Min 271.0 F 28.3
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 231.1 F 0.0
10:00 236.2 F 0.0
15:00 228.2 F 0.0
20:00 224.2 F 0.0
25:00 224.2 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

10/19/2023 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____