

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4622

Name: RITTER 1  
Sterilizer ID: V2178198  
06/10/2024 12:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 112.4 F 0.0  
2:00 113.0 F 1.0  
4:00 125.9 F 3.3  
6:00 150.7 F 6.1  
8:00 175.5 F 9.3  
10:00 198.8 F 13.3  
12:00 209.9 F 5.0  
14:00 218.7 F 8.9  
16:00 229.5 F 13.1  
18:00 232.5 F 8.8  
20:00 242.2 F 12.4  
22:00 241.9 F 11.8  
24:00 253.3 F 17.5  
26:00 263.6 F 23.5  
28:00 272.5 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.5 F 29.5  
2:00 271.9 F 28.9  
3:00 271.4 F 28.9  
4:00 272.4 F 29.5

Min 271.2 F 28.6  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.5 F 0.0  
5:00 201.1 F 0.0  
10:00 198.7 F 0.0  
15:00 198.8 F 0.0  
20:00 213.1 F 0.0  
25:00 218.4 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 28:18  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:06:58  
CYCLE COMPLETE

06/10/2024 01:53 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_