

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1743

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2022 01:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.6 F 0.0  
2:00 169.0 F 3.5  
4:00 193.6 F 8.0  
6:00 216.5 F 9.6  
8:00 227.9 F 8.8  
10:00 236.4 F 9.8  
12:00 243.0 F 12.5  
14:00 241.8 F 11.2  
16:00 255.9 F 18.5  
18:00 268.2 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.5  
1:00 272.0 F 28.7  
2:00 272.2 F 28.8  
3:00 272.0 F 28.8  
4:00 272.0 F 28.6  
  
Min 271.6 F 28.4  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.5 F 0.0  
5:00 229.8 F 0.0  
10:00 219.2 F 0.0  
15:00 217.4 F 0.0  
20:00 215.4 F 0.0  
25:00 211.0 F 0.0  
30:00 222.1 F 0.0  
  
Drying Complete

FILLING: 1:34  
HEATING: 19:03  
STERILIZING: 4:00  
VENTING: 2:54  
DRYING: 30:00  
TOTAL CYCLE: 00:57:30  
CYCLE COMPLETE

03/25/2022 07:15 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02