
Midmark M11 - v1.0.0
Total Cycles: 1825

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2022 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.2 F 0.0
2:00 198.5 F 3.3
4:00 195.6 F 8.3
6:00 214.4 F 9.6
8:00 229.5 F 9.1
10:00 237.4 F 9.4
12:00 243.1 F 12.4
14:00 241.6 F 11.1
16:00 255.3 F 18.1
18:00 267.2 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 271.9 F 28.6
2:00 272.2 F 28.6
3:00 272.6 F 28.9
4:00 272.5 F 29.2

Min 271.6 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 228.4 F 0.0
10:00 209.9 F 0.0
15:00 218.0 F 0.0
20:00 231.3 F 0.0
25:00 229.7 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:00
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:38
CYCLE COMPLETE

04/14/2022 12:25 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____