

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2052

Name: RITTER 1  
Sterilizer ID: V2178198  
06/24/2022 11:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
2:00 196.2 F 3.5  
4:00 208.8 F 7.7  
6:00 228.7 F 13.0  
8:00 228.2 F 7.8  
10:00 240.0 F 13.0  
12:00 238.6 F 10.1  
14:00 235.0 F 8.2  
16:00 249.4 F 14.8  
18:00 261.3 F 21.5  
20:00 272.1 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.1 F 28.8  
2:00 271.5 F 28.1  
3:00 271.4 F 28.5  
4:00 272.0 F 28.7  
  
Min 271.0 F 28.0  
Max 272.2 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.0 F 0.0  
5:00 224.3 F 0.0  
10:00 241.3 F 0.0  
15:00 239.7 F 0.0  
20:00 225.8 F 0.0  
25:00 228.1 F 0.0  
30:00 229.5 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 20:32  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:09  
CYCLE COMPLETE

06/24/2022 11:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_