

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5138

Name: RITTER 1  
Sterilizer ID: V2178198  
11/28/2024 10:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.8 F 0.0  
2:00 73.9 F 0.0  
4:00 90.0 F 0.0  
6:00 115.8 F 0.0  
8:00 153.7 F 0.1  
10:00 181.6 F 0.3  
12:00 205.3 F 0.5  
14:00 217.2 F 2.5  
16:00 225.2 F 5.5  
18:00 232.7 F 8.6  
20:00 242.1 F 12.5  
22:00 238.3 F 9.8  
24:00 238.4 F 9.3  
26:00 243.5 F 11.4  
28:00 246.6 F 13.3  
30:00 257.9 F 19.4  
32:00 268.0 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.1 F 28.5  
2:00 272.7 F 28.9  
3:00 272.7 F 28.7  
4:00 272.4 F 28.5  
  
Min 271.6 F 28.0  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
5:00 203.5 F 0.0  
10:00 220.0 F 0.0  
15:00 207.6 F 0.0  
20:00 206.6 F 0.0  
25:00 210.3 F 0.0  
30:00 210.5 F 0.0  
  
Drying Complete

FILLING: 0:41  
HEATING: 33:34  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:11:57  
CYCLE COMPLETE

11/28/2024 11:25 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_