

-----  
Midmark M11 - v1.0.5  
Total Cycles: 904

Name: RITTER2  
Sterilizer ID: V2313228  
03/25/2022 07:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.2 F 0.0  
2:00 84.9 F 0.0  
4:00 99.4 F 0.0  
6:00 124.5 F 0.0  
8:00 158.5 F 1.6  
10:00 175.1 F 4.8  
12:00 193.4 F 8.1  
14:00 207.6 F 11.8  
16:00 220.8 F 5.4  
18:00 232.0 F 9.4  
20:00 237.0 F 11.6  
22:00 239.8 F 11.1  
24:00 237.0 F 9.5  
26:00 250.1 F 15.8  
28:00 261.1 F 22.0  
30:00 271.3 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.5 F 29.3  
2:00 271.4 F 29.0  
3:00 272.4 F 29.3  
4:00 271.3 F 28.9

Min 271.2 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.5 F 0.0  
5:00 226.0 F 0.0  
10:00 189.1 F 0.0  
15:00 193.2 F 0.0  
20:00 203.3 F 0.0  
25:00 180.4 F 0.0  
30:00 172.1 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 30:40  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:08:51  
CYCLE COMPLETE

03/25/2022 08:25 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02