
Midmark M11 - v1.0.0
Total Cycles: 1364

Name: RITTER 1
Sterilizer ID: V2178198
11/11/2021 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.5 F 0.0
2:00 183.4 F 2.6
4:00 209.1 F 7.1
6:00 222.2 F 12.8
8:00 233.4 F 8.4
10:00 241.4 F 11.4
12:00 242.4 F 11.5
14:00 239.5 F 9.8
16:00 253.2 F 16.6
18:00 265.1 F 23.6
20:00 271.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 271.9 F 28.9
2:00 272.0 F 29.1
3:00 272.1 F 29.2
4:00 272.0 F 29.2

Min 271.8 F 28.6
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 229.8 F 0.0
10:00 224.5 F 0.0
15:00 229.1 F 0.0
20:00 228.7 F 0.0
25:00 225.7 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 20:26
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

11/11/2021 03:58 PM
ACCEPTED: X
REJECTED:
Operator: U02