
Midmark M11 - v1.0.5
Total Cycles: 1974

Name: RITTER2
Sterilizer ID: V2313228
01/25/2023 03:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.3 F 0.0
2:00 119.9 F 0.0
4:00 139.1 F 0.1
6:00 185.1 F 0.6
8:00 207.9 F 3.3
10:00 219.7 F 7.5
12:00 231.9 F 12.2
14:00 232.6 F 8.8
16:00 240.3 F 11.1
18:00 244.4 F 13.1
20:00 243.8 F 12.4
22:00 257.0 F 19.4
24:00 268.5 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.6 F 29.3
2:00 271.6 F 28.6
3:00 272.7 F 29.5
4:00 272.1 F 28.9

Min 271.6 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.3 F 0.0
5:00 222.4 F 0.0
10:00 222.3 F 0.0
15:00 220.9 F 0.0
20:00 221.8 F 0.0
25:00 228.5 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:12
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:03:21
CYCLE COMPLETE

01/25/2023 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____