
Midmark M11 - v1.0.5
Total Cycles: 1754

Name: RITTER2
Sterilizer ID: V2313228
11/21/2022 03:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.6 F 0.0
2:00 202.7 F 4.1
4:00 217.4 F 10.3
6:00 226.0 F 7.3
8:00 239.0 F 12.2
10:00 243.4 F 12.9
12:00 244.3 F 12.8
14:00 258.2 F 20.3
16:00 270.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.3 F 29.4
2:00 271.3 F 29.0
3:00 272.5 F 29.5
4:00 271.5 F 28.8

Min 271.2 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 229.4 F 0.0
10:00 243.6 F 0.0
15:00 235.1 F 0.0
20:00 227.2 F 0.0
25:00 222.4 F 0.0
Dry Cycle Aborted

FILLING: 2:25
HEATING: 16:50
STERILIZING: 4:00
VENTING: 3:11
DRYING: 26:47
TOTAL CYCLE: 00:53:13

11/21/2022 04:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____