

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1803

Name: RITTER2  
Sterilizer ID: V2313228  
12/05/2022 02:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.5 F 0.0  
2:00 169.2 F 3.0  
4:00 190.9 F 6.7  
6:00 213.8 F 11.7  
8:00 217.4 F 5.8  
10:00 229.9 F 11.0  
12:00 230.8 F 8.1  
14:00 240.4 F 11.3  
16:00 244.9 F 13.3  
18:00 257.4 F 20.1  
20:00 268.9 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.4  
1:00 271.8 F 29.3  
2:00 272.1 F 29.6  
3:00 272.4 F 30.0  
4:00 271.7 F 29.6  
  
Min 271.6 F 29.3  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.3 F 0.0  
5:00 220.4 F 0.0  
10:00 237.1 F 0.0  
15:00 226.8 F 0.0  
20:00 227.2 F 0.0  
25:00 228.9 F 0.0  
30:00 230.3 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 21:01  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:20  
CYCLE COMPLETE

12/05/2022 03:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_