
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002203
10:00 04/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:22
STERILIZING 4:29
136.5 C 224kPa 4:29
136.2 C 222kPa 4:59
135.9 C 220kPa 5:29
136.3 C 224kPa 5:59
136.4 C 225kPa 6:29
136.3 C 220kPa 6:59
136.1 C 222kPa 7:29
136.3 C 225kPa 7:59
136.3 C 225kPa 7:59
VENTING 7:59
AIR DRYING 8:51
CYCLE COMPLETE 9:55

Digital Signature #
8C7BF26E67E9A8CA
