

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002238  
11:24 11/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:47  
STERILIZING 3:51  
136.4 C 221kPa 3:51  
136.3 C 223kPa 4:21  
136.1 C 224kPa 4:51  
136.3 C 224kPa 5:21  
136.4 C 225kPa 5:51  
136.4 C 223kPa 6:21  
136.2 C 222kPa 6:51  
136.3 C 224kPa 7:21  
136.3 C 224kPa 7:21  
VENTING 7:21  
AIR DRYING 8:12  
CYCLE COMPLETE 24:24  
-----

Digital Signature #  
B34BCAE90149F319  
-----