
Midmark M11 - v1.0.0
Total Cycles: 4500

Name: RITTER 1
Sterilizer ID: V2178198
04/29/2024 08:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 74.2 F 0.1
4:00 110.0 F 0.7
6:00 126.4 F 3.2
8:00 154.4 F 5.9
10:00 168.1 F 8.5
12:00 189.6 F 11.9
14:00 203.1 F 3.2
16:00 216.1 F 6.3
18:00 229.4 F 9.8
20:00 237.2 F 10.8
22:00 235.6 F 9.6
24:00 240.6 F 9.8
26:00 245.5 F 12.9
28:00 257.1 F 19.0
30:00 267.6 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.7 F 28.8
2:00 272.0 F 28.5
3:00 272.0 F 28.2
4:00 272.7 F 28.9

Min 271.5 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.8 F 0.0
5:00 196.5 F 0.0
10:00 207.6 F 0.0
15:00 211.6 F 0.0
20:00 212.7 F 0.0
25:00 215.0 F 0.0
30:00 217.6 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:09:24
CYCLE COMPLETE

04/29/2024 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____