
Midmark M11 - v1.0.0
Total Cycles: 2465

Name: RITTER 1
Sterilizer ID: V2178198
10/20/2022 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.0 F 0.0
2:00 198.2 F 4.3
4:00 202.7 F 9.2
6:00 225.9 F 8.2
8:00 232.2 F 9.8
10:00 234.7 F 8.4
12:00 244.4 F 13.0
14:00 242.6 F 11.6
16:00 255.5 F 18.3
18:00 267.5 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 271.9 F 29.0
2:00 272.1 F 28.7
3:00 271.2 F 28.4
4:00 272.4 F 29.2

Min 271.2 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 231.2 F 0.0
10:00 227.7 F 0.0
15:00 226.5 F 0.0
20:00 227.5 F 0.0
25:00 230.9 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 2:39
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

10/20/2022 01:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____