

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4976

Name: RITTER 1  
Sterilizer ID: V2178198  
10/08/2024 10:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.0 F 0.0  
2:00 134.2 F 0.1  
4:00 141.9 F 0.6  
6:00 163.5 F 1.1  
8:00 184.1 F 4.6  
10:00 208.4 F 8.1  
12:00 218.6 F 12.0  
14:00 223.5 F 6.2  
16:00 233.9 F 11.0  
18:00 231.9 F 7.6  
20:00 245.2 F 13.5  
22:00 243.4 F 12.0  
24:00 255.3 F 18.1  
26:00 265.8 F 24.5  
28:00 271.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.7  
1:00 272.7 F 29.2  
2:00 271.8 F 28.6  
3:00 272.7 F 29.3  
4:00 272.2 F 28.8

Min 271.6 F 28.6  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
5:00 200.3 F 0.0  
10:00 196.0 F 0.0  
15:00 222.1 F 0.0  
20:00 236.0 F 0.0  
25:00 232.0 F 0.0  
30:00 226.1 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 28:37  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:07:11  
CYCLE COMPLETE

10/08/2024 11:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_