

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5025

Name: RITTER 1  
Sterilizer ID: V2178198  
10/25/2024 12:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 102.9 F 0.0  
2:00 107.9 F 0.1  
4:00 126.3 F 0.4  
6:00 158.2 F 0.8  
8:00 183.8 F 2.6  
10:00 198.2 F 5.7  
12:00 211.1 F 8.8  
14:00 221.3 F 12.3  
16:00 226.3 F 6.8  
18:00 236.5 F 11.5  
20:00 234.1 F 8.3  
22:00 243.4 F 11.9  
24:00 244.9 F 12.7  
26:00 256.8 F 18.8  
28:00 267.3 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.2 F 28.9  
2:00 272.5 F 28.7  
3:00 271.8 F 28.3  
4:00 271.2 F 28.1  
  
Min 271.2 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.5 F 0.0  
5:00 191.8 F 0.0  
10:00 195.0 F 0.0  
15:00 193.7 F 0.0  
20:00 193.0 F 0.0  
25:00 196.9 F 0.0  
30:00 211.7 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 29:37  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:08:14  
CYCLE COMPLETE

10/25/2024 01:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_