
Midmark M11 - v1.0.0
Total Cycles: 5427

Name: RITTER 1
Sterilizer ID: V2178198
03/20/2025 02:10 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.4 F 0.0
2:00 115.7 F 1.2
4:00 129.1 F 3.9
6:00 156.6 F 7.0
8:00 183.8 F 10.7
10:00 206.5 F 3.6
12:00 215.2 F 7.2
14:00 223.5 F 11.3
16:00 228.1 F 6.5
18:00 241.1 F 11.8
20:00 237.3 F 9.4
22:00 249.5 F 15.3
24:00 260.5 F 21.4
26:00 270.7 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.6 F 29.7
2:00 271.9 F 29.4
3:00 272.0 F 29.1
4:00 272.3 F 29.4

Min 270.9 F 28.5
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 206.0 F 0.0
10:00 222.8 F 0.0
15:00 215.0 F 0.0
20:00 216.7 F 0.0
25:00 220.5 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 26:42
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:05:14
CYCLE COMPLETE

03/20/2025 03:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____