
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001140
11:32 23/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:08
PRESSURIZING 3:16
STERILIZING 4:26
136.0 C 220kPa 4:26
136.1 C 223kPa 4:56
136.1 C 223kPa 5:26
136.1 C 222kPa 5:56
136.3 C 226kPa 6:26
136.3 C 224kPa 6:56
136.4 C 225kPa 7:26
136.4 C 225kPa 7:56
136.4 C 225kPa 7:56
VENTING 7:56
AIR DRYING 8:51
CYCLE COMPLETE 16:11

Digital Signature #
AFB6832A35594E00
